

Radiometer Medical ApS
Åkandevej 21
2700 Brønshøj
Denmark
Phone: +45 38 27 38 27
Fax: +45 38 27 27 27
CVR no. 27509185
www.radiometer.com

To whom it may concern

December 22, 2021

Apache Log4j Security Vulnerabilities

Dear valued customer

Apache Software Foundation has announced the discovery of the vulnerabilities named Log4j ([Log4j – Apache Log4j Security Vulnerabilities](#), <https://nvd.nist.gov/vuln/detail/CVE-2021-44228> and [NVD - CVE-2021-4104 \(nist.gov\)](#))

Radiometer Medical has evaluated whether these vulnerabilities have an impact on our products:

- ABL80
- ABL800
- ABL9
- ABL90
- AQT90
- TCM4
- TCM5
- AQUIRE*

It is our assessment that these vulnerabilities have no impact on all versions of the above devices. For this reason, we will not issue separate patches to the software used by Radiometer equipment.

This message constitutes an information service and is provided without guarantees and warranties of any kind, expressed or implied. Radiometer excludes any liability for actions or omissions performed based on this information.

Kind regards



Christian Knudsen
Technical Product manager

*Please note that Radiometer have investigated the Mirth versions recommended by Radiometer for use with AQUIRE, and not found these to be affected by the vulnerabilities reported. We recommend users of AQUIRE with Mirth installed to investigate that the versions they are using are the ones recommended by Radiometer, and if not take appropriate actions